



Member of the FM Global Group

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmaprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

This certificate is issued for the following equipment:

1616-5015. Wireless UT Thickness Node.
ISSUT5M-XX-YY-5015, with XX = S or TC, YY= 3 or 5. Transducer.
ISDUTXXM-5015, with XX = 2 or 5. Transducer

IS/I/1/ABCD/T4 for $-55^{\circ}\text{C} \leq T_a \leq +55^{\circ}\text{C}$; Control Drawing #1616-6000;
I/O/AEx ia/IIC/ T4 for $-55^{\circ}\text{C} \leq T_a \leq +55^{\circ}\text{C}$; Control Drawing #1616-6000;
IP66.

Special Conditions of Use:

- 1. Do not remove cover, replace batteries, fuse or plug-in module unless the area is known to be free of ignitable gas vapors. Only battery type MISTRAS E950-0036 shall be used.*
- 2. To avoid electrostatic discharge, cleaning must only be performed with a damp cloth.*
- 3. The enclosure contains aluminum and is considered a potential risk of ignition by impact or friction. Care must be taken during installation to prevent impact or friction.*

Equipment Ratings:

Intrinsically safe for Class I, Division 1, Groups A, B, C and D, Ex ia for Class I, Zone 0, Group IIC, temperature code T4 for $-55^{\circ}\text{C} \leq T_a \leq +55^{\circ}\text{C}$; IP66 hazardous locations.

To verify the availability of the Approved product, please refer to www.approvalguide.com

FM Approved for:

MISTRAS Group
Princeton Junction, NJ
USA

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

CSA C22.2 No. 61010.1	2012
CSA-C22.2 No. 60079-0	2011
CSA-C22.2 No. 60079-11	2014
CSA C22.2 No. 60529	2005 (R2015)


Original Project ID: 3051800

Approval Granted: February 29, 2016

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC



J. E. Marquedant
Manager, Electrical Systems

29 February 2016
Date

To verify the availability of the Approved product, please refer to www.approvalguide.com